

EC Certificate

Directive 93/42/EEC Annex II, excluding Section 4 Full Quality Assurance System Medical Devices

Registration No.: HD 60148266 0001

Report No.: 21253381 019

Manufacturer: RAPID Biomedical GmbH

Kettelerstr. 3-11 97222 Rimpar Deutschland

Products: RF coils for magnetic resonance systems

for medical diagnostics

Replaces Certificate, Registration No.: HD 60113456 0001

Expiry Date: 2024-05-26

The Notified Body hereby declares that the requirements of Annex II, excluding section 4 of the directive 93/42/EEC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex II, section 5 of the aforementioned directive. For placing on the market of class III devices covered by this certificate an EC design-examination certificate according to Annex II, section 4 is required.

Effective Date: 2020-04-23

Date: 2020-04-23

TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC concerning medical devices with the identification number 0197.

10/020 h 04.08 🕸 TÜV, TÜEV and TÜV are registered trademarks. Utilisation and application requires prior appro-